EXHIBIT C

```
Page 345
                 UNITED STATES DISTRICT COURT
                 FOR THE DISTRICT OF ARIZONA
In Re: Bard IVC Filters ) MD-15-02641-PHX-DGC
Products Liability Litigation )
                             ) Phoenix, Arizona
                             ) May 16, 2018
Doris Jones, an individual,
                             ) 1:00 p.m.
             Plaintiff,
                              ) CV 16-00782-PHX-DGC
        vs.
C.R. Bard, Inc., a New
Jersey corporation; and Bard )
Peripheral Vascular, Inc., an )
Arizona corporation,
             Defendants.
      BEFORE: THE HONORABLE DAVID G. CAMPBELL, JUDGE
             REPORTER'S TRANSCRIPT OF PROCEEDINGS
             (Jury Trial - Day 2 - P.M. Session)
              (Pages 345 through 468, inclusive.)
Official Court Reporter:
Laurie A. Adams, RMR, CRR
Sandra Day O'Connor U.S. Courthouse, Suite 312
401 West Washington Street, Spc 43
Phoenix, Arizona 85003-2151
(602) 322-7256
```

Proceedings Reported by Stenographic Court Reporter Transcript Prepared by Computer-Aided Transcription

Page 403 the -- or fatigue of the arms and legs themselves? 1 2 A. Well, yeah, the arms and legs are designed so that they are very susceptible to fatigue. And that's tied up with their 3 4 shape and orientation and the fact that they can perforate the wall of the IVC. And that leads to enhanced likelihood of 5 02:29PM 6 fracture by fatigue. 7 MR. STOLLER: Gay, would you find Exhibit 4559, 8 please. 9 THE COURT: We're going to break at this point, Mr. 10 Stoller. We will take a break until 2:45. Ladies and 02:30PM 11 Gentlemen, we'll excuse you at this time. 12 MR. STOLLER: Thank you, Your Honor. 13 (Recess from 2:30 p.m. until 2:47 p.m.) THE COURT: Go ahead, Mr. Stoller. 14 15 MR. STOLLER: Thank you, Your Honor. 02:47PM BY MR. STOLLER: 16 17 Q. Dr. McMeeking, just before we broke we were talking about 18 your opinions with respect to the defective -- the design defects in the Eclipse Filter. And you identified the conical 19 20 design, the thin needle-like limbs in the arms and legs that 02:47PM 21 lead to perforation, the design concentration -- the design 22 that concentrates the strain mentioned at the shoulder or the 23 elbow where the filter meets the cap. 24 Are there any alternative design elements that you 02:47PM 25 have identified that would help relieve or ameliorate some of

```
Page 404
     the issues?
 1
 2
             MR. NORTH: Your Honor, objection. That's outside the
 3
     scope of the report.
 4
             THE COURT: Is that in the report, Mr. Stoller?
             MR. STOLLER: It is in the report.
                                                                     02:47PM
 5
 6
             THE COURT: All right. Could I have a copy and have
 7
    you show that to me? Why don't you approach, counsel, and we
 8
     can talk about that.
 9
             Ladies and Gentlemen, if you want to stand up feel
10
     free.
                                                                     02:48PM
              (Discussion was had at sidebar out of the hearing of
11
     the jury:)
12
13
             THE COURT: By the way, while we're looking for that,
14
     the jurors, while they were coming back in, asked Traci if they
15
     could hold the filter, if it could be passed around so they
                                                                     02:48PM
     could each hold it and look at it.
16
17
             MR. STOLLER: This is Dr. McMeeking's filter so we
18
    will have to ask him.
             MR. NORTH: It has to be admitted into evidence,
19
20
    doesn't it?
                                                                     02:48PM
21
             MR. STOLLER: No.
22
             THE COURT: Do they want to pull the leg off?
23
             MR. STOLLER: I'm not sure I mind if they pull the leg
24
          I think he might. It's fine with us. I will have Dr.
    McMeeking give it to them and they can pass it around.
25
                                                                     02:48PM
```

```
Page 405
             THE COURT: So where are we?
1
 2
             MR. STOLLER: Question was alternative design
 3
     elements. One is, he identified a rounded edge of the cap as
 4
     fixing the issue at least on Page 62 of his report.
             THE COURT: Anybody have a copy?
                                                                    02:48PM
 5
 6
             MR. STOLLER: I do. Got 80 something pages.
 7
             MS. HELM: The report is 25 pages.
 8
             MR. STOLLER: This is the March 3, 2017 report. You
9
    are looking at one of the reports from the Barraza case. He
10
     talks about how you can change it to make it better, sharp
                                                                    02:49PM
11
     corner --
             THE COURT: Hold on. Well, he's saying a sharp corner
12
13
     is bad.
14
             MR. STOLLER: He's saying a rounded corner --
15
             THE COURT: Where does it say that?
                                                                    02:49PM
16
             MR. STOLLER: It's this one, Your Honor. 12.
17
    Possibility of a strain concentration.
18
             THE COURT: All right. I see that. Do you have a
    response to that?
19
20
             MR. NORTH: The demonstrative we have been shown that 02:50PM
21
     they intend to use goes far beyond those elements. And in his
22
    deposition he twice disclaimed any alternative design.
23
             THE COURT: So what -- so this is simply talking
24
     about --
25
            MR. STOLLER: Talking about --
                                                                    02:50PM
```

```
Page 406
             THE COURT: Hold on.
 1
 2
             MR. STOLLER: Sorry, Your Honor.
 3
             THE COURT: By curving or chamfering the edge or by
 4
     eliminating the connection to the cap. Okay. So that's in
     there. What else do you have in the way of alternative design? 02:51PM
 5
 6
    Well, let me ask you this. Are you going to elicit alternative
 7
    design testimony besides that?
 8
             MR. STOLLER: Yes.
 9
             THE COURT: What else are you going to elicit?
10
             MR. STOLLER: Based on the examination where he was
11
     examined by Ms. Daly we're going to elicit testimony that they
     could have had caudal anchors to elicit caudal movement.
12
13
             THE COURT: Where is that?
             MR. STOLLER: Page 32. The July 6, 2017: Are there
14
15
    any other changes you think could be made to the filters? He
                                                                    02:51PM
    discusses caudal anchors and penetration limiters.
16
17
             THE COURT: All right. I see that on Page 32. What
18
     else are you going to elicit?
             MR. STOLLER: He also testified, Your Honor, at the
19
20
     last trial that the two tier -- the two tier staging of the SNF 02:52PM
21
     is a better design to alleviate tilt.
22
             THE COURT: What is your response on those three
23
    points.
             MR. NORTH: Your Honor, I think I'm just going to --
24
     I'm sorry. He says the opposite in other places. I guess I
25
                                                                    02:52PM
```

```
Page 407
     just need to impeach him with that.
 1
 2
             THE COURT: So the three areas are where the arms and
 3
     legs leave the cap; the angle while attaching to the cap,
 4
     caudal anchors, penetration limiters.
             MR. STOLLER: Those are two separate things.
                                                             02:52PM
 5
 6
             THE COURT: I know. And the two-staged tiered
 7
     approach of the Simon Nitinol. That's what you're eliciting?
 8
             MR. NORTH: Is the two-tiered fair game just because
 9
     it was mentioned in the last trial?
10
             THE COURT: Well, when I said how do you feel about
                                                                    02:53PM
11
     that, you didn't state an objection.
             MR. NORTH: I'm asking.
12
13
             THE COURT: Are you asking if I would advise you to
    object?
14
15
             MR. NORTH: Fair enough. We would object because it
                                                                    02:53PM
16
    was not explored in either the deposition or the --
17
             THE COURT: Is the two-tiered in the deposition or
18
    report?
19
             MR. STOLLER: He identified the SNF as all ulterior
20
    design in his report.
                                                                    02:53PM
21
             THE COURT: Can you show me where?
22
             MR. STOLLER: Your Honor, I will skip that one in the
23
     interest of time. I know I'm not going to deal with it.
             THE COURT: Okay.
24
                                                                    02:53PM
25
             (In open court.)
```

Page 408

- 1 THE COURT: Thanks, Ladies and Gentlemen.
- 2 BY MR. STOLLER:
- 3 Q. Dr. McMeeking, I want to talk about some potential
- 4 alternative design features that you believe may have fixed
- 5 some of the issues you have identified. Let me talk 02:54PM
- 6 specifically about the issues you identified about the IVC
- 7 filter leg coming out of the cap and the concentration of
- 8 strains there. Do you understand what I'm talking about?
- 9 A. Yes. Yes, I do.
- 10 Q. What sort of design changes could Bard have made that would 02:54PM
- 11 have fixed or reduced the problems with that design issue?
- 12 A. They could have smoothed the sharp corner to make it
- 13 gentler.
- 14 Q. What effect would that have had?
- 15 A. That would have the effect of avoiding the severe 02:54PM
- 16 concentration of strain that can occur at that location.
- 17 Q. And in turn result in less fractures?
- 18 A. Correct.
- 19 Q. Okay. With respect to the issue of perforation, what are
- 20 there -- do you have opinions as to what sort of design changes 02:55PM
- 21 they could have made to either eliminate or reduce the
- 22 frequency of that occurrence?
- 23 A. Yes. They could have added penetration limiters which
- 24 would be features on the limbs that would help to slow down or
- 25 $\,$ even stop the perforation of the limbs through the wall of the $\,$ 02:55PM $\,$

Page 409

- 1 IVC.
- 2 Q. Does that have effect on migration as well?
- 3 A. If designed properly, they would have helped to eliminate
- 4 migration down towards the feet. The hooks which are on the
- 5 filter already help to limit the tendency for the filter to 02:55PM
- 6 move toward your head. But adding some hook-like features that
- 7 would point in the opposite direction would help to stop the
- 8 filter going downwards.
- 9 Q. And is there a difference between a design element that you
- 10 would have added to alleviate the effect of movement down 02:56PM
- 11 versus the perforation?
- 12 A. Well, they could have been the same device. But in one
- 13 case it could have -- hook-like features would have worked. In
- 14 the other case of penetration limitation, somewhat blunt
- 15 features would help to slow down the motion of the limb through 02:56PM
- 16 the wall of the IVC.
- 17 Q. And the migration downward here, what the jury has been
- 18 told is called caudal?
- 19 A. Caudal migration.
- 20 Q. You would put on something called caudal anchors? 02:56PM
- 21 A. They are called caudal anchors.
- 22 Q. Where would those be placed?
- 23 A. They could be placed either at the feet or the hands of the
- 24 limbs, and they may be placed elsewhere on the limbs as is
- 25 convenient for the way that the device is hooked into the IVC 02:56PM

Page 410

- 1 and the way that it's delivered from the delivery tube.
- 2 Q. Let's talk for a moment, Doctor, about your opinions with
- 3 respect to Bard's design processes and testing for the Eclipse
- 4 Filter in the prior generation. And I believe you testified
- 5 earlier that you have opinions with respect to Bard's actions 02:57PM
- 6 in testing its filters. What is your opinion?
- 7 A. My opinion is that the testing was inadequate.
- 8 Q. And why do you hold that opinion?
- 9 A. I hold that opinion because in some of the tests that they
- 10 did do the conditions that were used were not the worst-case 02:57PM
- 11 conditions that were reasonably foreseeable. And in other
- 12 cases, there was no test done at all. For example, there was
- 13 no test to look at whether the filter would tilt after it had
- 14 been implanted. And there was no test to look at whether the
- 15 limbs would have a tendency to cut through the wall of the IVC 02:57PM
- 16 and, therefore, puncture and perforate through the wall of the
- 17 IVC.
- 18 Q. Did you review Bard's bench testing?
- 19 A. I did.
- 20 Q. Did you review its finite element analysis?

02:58PM

- 21 A. I did.
- 22 Q. Was its testing reasonable?
- 23 A. No.
- 24 Q. Was its testing adequate?
- 25 A. No. It wasn't adequate. No.

02:58PM